

Calibration Certificate

Certificate No.:	TZ2022052302	Certificate Date:	2022/5/23
-------------------------	--------------	--------------------------	-----------

Reference Instrumentation

Device	Manufacturer	Equipment No.	Accuracy	Calibration Date
Programmable Temperature & Humidity Chamber	Donguan Kesionots Precision Instrument	KS-HW150	±0.05°C; ±3%RH	August, 23rd, 2021
High Accuracy RTD Thermometer	OMEGA	HH376	±0.15°C	August, 23rd, 2021

Tzone Digital Technology Co., Ltd. certifies that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. These device were assembled, tested and calibrated in accordance with its specifications and prior to release for shipment. All reference instruments are calibrated by SHENZHEN HSR MEASUREMENT AND INSPECTION CO.,LTD, and Shenzhen Zhongzhun Testing Co.,Ltd. a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according Mutual Recognition Agreement.

Product Information

Product Name	Model	Serial No.	Production Date	Expiry Date	Quantity
TZONE PDF USB TEMPERATURE DATA LOGGER	TempU02	TZ2203012051-TZ2203012100	2022/5/23	2024/5/22	50PCS

Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

Validation Points	Tolerance	Result
-20°C	±0.5°C	Pass
0°C	±0.5°C	Pass
30°C	±0.5°C	Pass

Test Result Pass/Fail	Pass	Signature of Calibration Technician	<i>Andy Zhang</i>
------------------------------	------	--------------------------------------------	-------------------

TZONE DIGITAL TECHNOLOGY COMPANY LIMITED

16D, Haiying building, South of Caitian Road, Futian District, Shenzhen City, China, 518033

Tel: 0086-755-82840647 Fax: 0086-755-82840648 Email: sales@tzonedigital.com Web: www.tzonedigital.com

For and on behalf of
TZONE DIGITAL TECHNOLOGY COMPANY LIMITED
天圖數碼科技有限公司

.....
Authorized Signature(s)