


PESSL INSTRUMENTS *GES.M.B.H.*

Werksweg 107, A-8160 WEIZ, AUSTRIA

Tel.: ++(0043) 3172 5521, Fax.: ++(0043) 3172 5521 23

E-Mail: sales@metos.at Webpage: <http://www.metos.at>

SENSOR	Pyranometer Sensor
Art.No.	IM506D



PARAMETER	
Sensor	LI-200SZ
Calibration	Calibration against Kipp and Zone CMP3 under daylight. Absolute error max. 5%, typically 3%
Stability	2% drift on 2 years use
Time to measure	10µs
Temperature dependency	0.15% per Centigrade
Cosines correction	Sensor corrects up 80 ° degrees
Direction error	1% through 360 degrees at 45°
Working temperature	-20 to 65 C
Relative Humidity	0 to 100%
Sensor	Photodiode
Housing	Weather resistance in stainless steel and aluminium
Size	12.68 cm lengths x 4.50 cm height
Weight	114g
Evaluation	Pulse Wide Modulation 0-80%=0-2000 W/m2